

ECG Contest

Response to ECG, October 2015



Respuesta al ECG de octubre de 2015

Óscar Cano

Unidad de Arritmias, Área de Enfermedades Cardiovasculares, Hospital Universitari i Politècnic La Fe, Valencia, Spain

During the patient's spontaneous rhythm, there are failures in atrial capture, indicated by the presence of occasional atrial spikes after atrial depolarization has taken place. Under the effect of the magnet, the atrial spike is seen to depolarize the ventricle because the atrial lead has become displaced and has descended to the ventricle, as shown in the radiological image (Figure).




Figure.

SEE RELATED ARTICLE:

<http://dx.doi.org/10.1016/j.rec.2015.05.004>

E-mail address: cano_osc@gva.es

<http://dx.doi.org/10.1016/j.rec.2015.05.005>

1885-5857/© 2015 Sociedad Española de Cardiología. Published by Elsevier España, S.L.U. All rights reserved.