

Minimally-invasive Transesophageal Echocardiography for Left Atrial Appendage Occlusion With a Latest-generation Microprobe. Initial Experience



Ecocardiografía transesofágica mínimamente invasiva con microsonda de última generación para el cierre percutáneo de la orejuela izquierda. Experiencia inicial

To the Editor,

Echocardiography imaging guidance plays an essential role in the safe and accurate execution of transcatheter structural interventions. Transesophageal echocardiography (TEE) is the preferred option in the guidance of left atrial appendage occlusion (LAAO) and transcatheter valve interventions.¹ Procedures with general anesthesia increase the risk of hypotensive periods, respiratory disorders, and the time and cost of interventional procedures. In elderly patients, general anesthesia is associated with perioperative pulmonary complications, neuromuscular block, and a higher risk of cognitive decline.² In addition, the use of a standard TEE probes for more than 60 minutes has been associated with a higher risk of oropharyngeal lesions.³

Currently, a large number of structural procedures are performed under local rather than general anesthesia. In prolonged procedures, however, TEE imaging is generally performed under general anesthesia, as standard TEE probes may not be very well tolerated with just light sedation. Intracardiac echocardiography (ICE) has been proposed as a valid alternative to TEE for some structural procedures such as LAAO to avoid general anesthesia. Despite the lower imaging quality, the higher cost and the need for additional vascular and transeptal access, some operators have moved to this less invasive approach. The recent arrival of novel micro-TEE probes might represent a very valid alternative to allow a minimally-invasive approach while maintaining proper imaging quality. Although some other experiences with micro-TEE probes have already been published,^{4,5} the current report presents the initial experience with a latest-generation micro-TEE probe (10T-D General electrics), which allows multiplane 2-dimensional (2D) imaging with a very small distal tip diameter of 5.6 × 7.7 mm.

We present our initial experience with 2 cases of LAAO in patients with atrial fibrillation with a high risk of stroke and with contraindications to long-term anticoagulation. The procedures were performed under light sedation and local anesthesia guided with a new-generation micro-TEE probe together with angiography. The micro-TEE probe provided excellent imaging quality for transeptal puncture and LAA assessment in both cases compared with the standard 2D-TEE, which was performed 1 week before the procedure (Figure). A 22-mm and 28-mm Amulet (St Jude Medical) were successfully implanted without complications. A final angiogram confirmed the correct placement of the device. The procedures lasted 52 and 45 minutes, respectively, and the patients were discharged the day after the procedure.

LAAO requires transeptal puncture, accurate LAA assessment and continuous imaging guidance for device deployment. The patient must be immobile during all these phases, as a sudden movement can increase the risk of significant complications, including atrial perforation and pericardial tamponade. For these reasons, most operators using TEE guidance tend to use general anesthesia.⁶ In our series, the micro-TEE probe was well tolerated by patients, and provided sufficient image quality for procedure guidance.

A previous study comparing micro-TEE with ICE and standard TEE showed that micro-TEE offers comparable image quality with a much wider field of vision than ICE and a slightly reduced but sufficient image quality and comparable accuracy to standard TEE.⁵ One of the main limitations of current micro-TEE probes and ICE is the absence of 3D imaging. For this reason, before LAAO, previous imaging assessment with 3-dimensional (3D) techniques (3D-TEE or computed tomography scan) is generally recommended. It is well established that 2D imaging might be sufficient to guide some structural procedures but definitely not enough to provide an accurate morphological assessment of cardiac structures like LAA. Structural procedures are the most demanding and 3D-TEE and simultaneous orthogonal views are very important during the procedure and if there are complications.

Although this initial experience focused on LAAO, the high patient-comfort and the excellent imaging quality suggested that




Figure. Comparison between adult and micro 2D-TEE probe. 2D-TEE adult (A) and micro (B) probes. Multiple 2D-TEE views with the adult (C,E,G) and micro probe (D,F,H). Final result after LAAO with fluoroscopy (I) and micro-TEE probe (J). LAAO, left atrial appendage occlusion; TEE, transesophageal echocardiographic.

latest-generation micro-TEE probes might represent a very valid alternative not only for LAO but also for other percutaneous structural heart interventions and could be an option in patients without intubation or general anesthesia, in particular in elderly patients or those with severe comorbidities. Additional experience will be necessary to confirm that high quality imaging is also observed in a larger sample of patients.

Gustavo Jiménez Brítez,* Laura Sanchis, Ander Regueiro, Manel Sabate, Marta Sitges, and Xavier Freixa

Instituto Cardiovascular, Hospital Clínic de Barcelona, Institut d'Investigacions Biomèdiques August Pi i Sunyer (IDIBAPS), Barcelona, Spain

* Corresponding author:

E-mail address: freixa@clinic.ub.es (G. Jiménez Brítez).

Available online 23 June 2018

REFERENCES

1. Wunderlich NC, Beigel R, Swaans MJ, Ho SY, Siegel RG. Percutaneous interventions for left atrial appendage exclusion: options assessment, and imaging using 2D and 3D echocardiography. *JACC Cardiovasc Imaging*. 2015;8:472–488.
2. Zangrillo A, Mazzone P, Votta CD, Villari N, Della Bella P, Monaco F. Prolonged transesophageal echocardiography during percutaneous closure of the left atrial appendage without general anesthesia: the utility of the Janus Mask. *Can J Anesth*. 2016;63:962–965.
3. Barreiro-Pérez M, Li CH, Mesa D, Fernandez-Golfin C, López-Fernández T, Pérez de Isla L. Selection of the Best of 2017 in Cardiac Imaging and Structural Interventionism. *Rev Esp Cardiol*. 2018;71:125–126.
4. Toole BJ, Slenick TC, Kreeger J. The Miniaturized multiplane Micro-Transesophageal echocardiography: A comparative evaluation of its accuracy and image quality. *J Am Soc Echocardiogr*. 2015;28:802–807.
5. Nijenhuis VJ, Alipour A, Wunderlich NC, et al. Feasibility of multiplane microtransesophageal echocardiographic guidance in structural heart disease transcatheter interventions in adults. *Neth Heart J*. 2017;25:669–674.
6. Freixa X, Martín V. Cierre percutáneo de orejuela izquierda. *Rev Esp Cardiol*. 2013;66:919–922.

<https://doi.org/10.1016/j.rec.2018.05.016>

1885-5857/

© 2018 Sociedad Española de Cardiología. Published by Elsevier España, S.L.U. All rights reserved.

Three-dimensional Image Fusion of Precatheter CT and MRI Facilitates Stent Implantation in Congenital Heart Defects



La fusión de imágenes tridimensionales por TAC y RM antes del cateterismo facilita el implante de stent en defectos cardíacos congénitos

To the Editor,

Advances in 3-dimensional (3D) image fusion software have enabled easy application of preintervention imaging, including computed tomography (CT) and magnetic resonance imaging (MRI) scans, to create a reliable roadmap for swift manipulation through complex cardiac anatomy.^{1,2} In an article published in *Revista Española de Cardiología*, Sandoval et al.³ presented an elegant case report of image fusion of a CT-derived 3D roadmap for aortic stent implantation.

We performed a retrospective review of stent implantation guided by novel image fusion software (VesselNavigator, Philips Healthcare) during the first 9 months since its introduction (Figure). Patient characteristics and catheterization data were reviewed with a focus on the fusion of preintervention imaging and intervention guidance, including structures used for registration, accuracy of 3D reconstruction overlay, and the need for intraprocedural roadmap readjustment.

VesselNavigator was used in all consecutive patients with available precatheterization CT or MRI scans. The 4-step workflow with VesselNavigator, including segmentation, planning, registration and live guidance, has been described elsewhere.^{4,5} For the registration, stored fluoroscopy images in 2 projections were fused with the 3D roadmap (2D-3D registration) using a combination of internal markers such as bony structures, the cardiac silhouette, areas of calcification, previously implanted devices, and vessel borders visualized with injection of a small amount of contrast medium.

Eighteen patients underwent transcatheter stent implantation for the creation of a “landing zone” prior to percutaneous pulmonary valve implantation (n = 7), pulmonary artery stenosis (n = 5), aortic coarctation (n = 3; [video of the supplementary material](#)), aortopulmonary collateral (n = 2) and arterial duct (n = 1) (Table). A 3D roadmap was created from existing CT (n = 17) or MRI (n = 1) datasets. For registration, the spine and vertebrae (n = 17; 94%), calcifications (n = 7, 39%), test angiogram (n = 5, 28%) or a previously placed device (n = 4, 22%) served as reference points for orientation of the 3D roadmap against live fluoroscopy. Accurate overlay was achieved in all but 2 patients (11%) with an additional 2 patients (11%) requiring intraprocedural realignment due to the introduction of a stiff wire and balloon/stent assembly. All stents were deployed at desired locations with no complications. In 13 patients (72%), the stent was positioned and deployed without prior contrast injection.

To date and to our knowledge, this is the first series focusing on application of preregistered 3D datasets exclusively for guidance of stent implantation in congenital heart defects. Our initial experience shows that, with modern fusion software, noninvasive 3D images may be easily and effectively reused for positioning and implantation of stents to various locations.^{2,4,5} A simple registration protocol requires only stored fluoroscopy in 2 projections, akin to the typical setup of isocenter at the beginning of each heart catheterization, and uses a combination of anatomical or iatrogenic landmarks or the injection of a small contrast volume. This 2D-3D registration enabled a shortening of the duration of interventional procedures and a reduction in contrast and radiation exposure.⁶ In several more recent patients, we felt confident enough to perform stent implantation without prior contrast injection.

To test the reproducibility of these promising individual results, we have initiated a prospective, international registry with contributions from Berlin (Germany), Sejong (South Korea), Mexico City (Mexico), Denver (United States), and Lodz (Poland).